

## **GEES QUALITY STANDARD REMARKS FOR MRO**

### **REMARK #**

### **SUBJECT / TEXT**

#### **ACSC-QC-021 (Rev. B, 8/11/00) CERTIFICATE OF CALIBRATION SERVICES**

AN ORIGINAL CERTIFICATE OF CONFORMANCE IS REQUIRED FOR MEASUREMENT AND TEST EQUIPMENT, MEASUREMENTS STANDARDS AND/OR CALIBRATION SERVICES PROVIDED UNDER THIS PURCHASE ORDER. CALIBRATION MUST BE PERFORMED UTILIZING MEASUREMENT STANDARDS WHOSE CALIBRATION SOURCE IS TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (N.I.S.T.) OR EQUIVALENT NATIONAL STANDARD. THE CERTIFICATE OF CONFORMANCE MUST INCLUDE AS A MINIMUM:

- 1) IDENTIFICATION OF THE CALIBRATION ITEM.
- 2) CALIBRATION SOURCE TRACEABLE TO N.I.S.T. OR EQUIVALENT.
- 3) DATE OF CALIBRATION
- 4) CALIBRATION RESULTS AND ASSIGNED VALUE.
- 5) ENVIRONMENTAL CONDITIONS UNDER WHICH THE RESULTS WERE OBTAINED
- 6) STATEMENT OF UNCERTAINTY AND LEVEL OF CONFIDENCE.